



**OREGON STATE UNIVERSITY  
REQUEST FOR QUOTE (RFQ)**

		<b>ISSUE DATE:</b>	September 1, 2015
<b>RFQ #</b>	SS177491RFQ	<b>RFQ DUE DATE:</b>	September 22, 2015 2:00PM PST
<b>DELIVER TO:</b>		<b>REQUESTED BY / RETURN QUOTE TO:</b>	
<b>DEPARTMENT:</b>	College of Agricultural Sciences	<b>NAME:</b>	Shoshana Shabazz
<b>ADDRESS:</b>	Gilmore Hall, 116 Gilmore Hall	<b>E-MAIL:</b>	<a href="mailto:Shoshana.Shabazz@oregonstate.edu">Shoshana.Shabazz@oregonstate.edu</a>
<b>CITY, STATE ZIP:</b>	Corvallis, OR, 97331	<b>TELEPHONE:</b>	541-737-0922
<b>REQUIRED DELIVERY DATE:</b>	November 30, 2015	<b>FAX:</b>	541-737-4810

Oregon State University (OSU) is seeking responsive, responsible quotes to **provide and install** a steam sterilizer (autoclave). In order to qualify as responsive you must meet the following specifications and qualifications:

Specifications:

- Materials:
  - Steam sterilizer (autoclave)
- Services:
  - Installation
    - Exact installation locations to be determined after award. Coordination for installation and equipment location to be done directly with College of Agricultural Sciences authorized personnel.

**In response to this RFQ, interested parties shall submit the following:**

- 1. The Quote Price Form on the following page, fully completed and signed.**

**DO NOT SUBMIT ANY OTHER DOCUMENTS INCLUDING BUT NOT LIMITED TO, A QUOTE ON COMPANY LETTERHEAD, COMPANY TERMS AND CONDITIONS OR PRODUCT BROCHURE. ONLY SUBMIT THE REQUIRED SUBMITTALS LISTED ABOVE. ANY DOCUMENTS SUBMITTED OTHER THAN THE REQUIRED SUBMITTALS LISTED ABOVE MAY CONSTITUTE CAUSE FOR QUOTE REJECTION.**

**OSU'S TERMS AND CONDITIONS GOVERNING THE PURCHASE RESULTING FROM THIS RFQ ARE INCLUDED IN EXHIBIT A AND WILL PREVAIL.**

*Remainder of page intentionally left blank.*



**OREGON STATE UNIVERSITY  
REQUEST FOR QUOTE (RFQ)**

**SS177491RFQ – STEAM STERILIZER (AUTOCLAVE)  
QUOTE PRICE FORM**

<b>Installer Name and Valid Oregon CCB Number</b>	<b>Name:</b>
	<b>Oregon CCB Number:</b>
<b>ITEM</b>	<b>PRICE</b>
Steam Sterilizer (Autoclave)	\$
Installation	\$

**Shipping, freight and handling must be included in quoted prices. Additional costs for such are disallowed.**

<b>TOTAL</b>	
--------------	--

<b>DELIVERY TIME AFTER RECEIPT OF ORDER:</b>		<b>PRICES VALID THROUGH:</b>	
--	--	------------------------------	--

<b>SPECIAL INSTRUCTIONS:</b>	<b>VENDOR INFORMATION:</b>	
<p>1. Unless otherwise specified, all items quoted are to be new, unused and not remanufactured in any way.</p> <p>2. Brand names are for the purpose of describing and establishing the characteristics desired and are not intended to limit or restrict competition. Interested firms may submit quotes for substantially equivalent products unless the RFQ provides that a specific brand is necessary because of compatibility requirements, etc. All such brand substitutions shall be subject to approval by OSU.</p> <p>3. You must clearly identify all products quoted. Brand name and model or number must be shown.</p> <p>4. Only documents issued as addenda by OSU serve to change the RFQ in any way.</p> <p>5. OSU reserves the right to make the award by item, partial or whole lots, groups of items or entire quote, whichever is in the best interest of OSU.</p> <p>6. OSU may reject any Quote not in compliance with the RFQ, attachments, and addenda, or if it is in the best interest of OSU.</p>	<b>COMPANY:</b>	
	<b>ADDRESS:</b>	
	<b>CITY, STATE, ZIP:</b>	
	<b>CONTACT NAME:</b>	
	<b>E-MAIL:</b>	
	<b>TELEPHONE:</b>	
	<b>FAX:</b>	
	<b>VENDOR SIGNATURE:</b>	
	<i>By signature below the undersigned certifies that they are authorized to act on behalf of the quoter and will comply with all aspects of the quote herein.</i>	
	<b>SIGNATURE:</b>	
	<b>NAME/TITLE:</b>	

This procurement is subject to the indicated Oregon State University Standard Terms and Conditions for:  Goods  Services  Purchase Order Construction  Software. The indicated terms and conditions may be viewed at <http://pacs.oregonstate.edu/terms-and-conditions>



# OREGON STATE UNIVERSITY REQUEST FOR QUOTE (RFQ)

## EXHIBIT A TERMS AND CONDITIONS

The following Terms and Conditions are in addition to those included on the following pages:

- Compliance with Bureau of Labor and Industries
  - Labor provided on site to support this RFQ falls under the Bureau of Labor Industries guidelines and as such, along with the Terms and Conditions located further in Exhibit A, the following is incorporated into this RFQ. **Installation MUST be performed by a commercially licensed contractor or subcontractor.** This license must be issued by the Oregon Construction Contractors Board prior to submitting a quote. Contractor and all subcontractors shall comply with the provisions of ORS 279C.800 through 279C.870, relative to Prevailing Wage Rates as outlined in Sections C.1 and C.2 of the General Conditions. This Solicitation and the resulting Purchase Order are subject to the following BOLI wage rate requirements, which are incorporated herein by Exhibit C.

These BOLI wage rates are available on line at: [http://www.boli.state.or.us/BOLI/WHD/PWR/pwr\\_state.shtml](http://www.boli.state.or.us/BOLI/WHD/PWR/pwr_state.shtml).

- Insurance Certificates and associated Endorsements will be required after award and prior to issuance of purchase order. Specifically required are
  - General Liability and associated Endorsement
  - Auto Liability and associated Endorsement
  - Workers Compensation

Include as additional insured, by endorsement, on the liability policies, " Oregon State University, its officers, board members, agents and employees." **Provide the endorsement w/policy # attached.** Insurance coverages required under this Contract shall be obtained from acceptable insurance companies or entities authorized to do business in the State of Oregon. The Contractor shall be financially responsible for all deductibles, self-insured retentions and/or self-insurance included hereunder. See Exhibit D

**Note: Insurance documents must come directly from your agent by either email (PDF) or standard mail. Faxed copies or documents delivered by the contractor/vendor, unless in a sealed envelope from your agent, will not be accepted. Insurance documents must be provided to OSU prior to the execution of a purchase order.**

*Remainder of page intentionally left blank.*

**EXHIBIT B  
EQUIPMENT SPECIFICATIONS**

Minimum specifications for steam sterilizer ( autoclave) to be located in the Biological & Ecological Engineering Department, Gilmore Hall, Oregon State University:

1. The bid is for new or remanufactured units. **Carefully study items 7 and 9 if unit is remanufactured.**
2. The autoclave may be a cabinet or stand-alone installation, or suitable for counter top operation. The chamber can be in vertical or horizontal configuration.
3. The interior chamber size should be rectangular and at least 19" wide x 19" high x 38" deep. The chamber may also be cylindrical in shape as long as it meets the minimum size dimensions specified. There should be a single door. The chamber gaskets should be silicon rubber or similar material and easily replaceable without tools.
4. The autoclave should operate at a temperature of 122°C or greater and be capable of vacuum, gravity, and slow exhaust (liquid) operation. Ancillary equipment (air compressors, etc.) necessary for operation in these modes must be included as part of the bid.
5. The autoclave should be fitted with microprocessor controls that control all aspects of its operation. In addition, a removable rack and at least two removable shelves should be supplied. Both the rack and shelves should be made of a corrosion resistant material.
6. The autoclave should be supplied with an integral electric steam generator. The autoclave facility has a 208 V, 3 phase and 115 V single phase electrical supply. The integral electric steam generator should be 208 V, 3 phase. The other electrical options should be capable of operating with the facility electric supply.
7. For a remanufactured unit, the unit must have been cleaned, disinfected, completely disassembled and inspected during the remanufacturing process. The pressure vessel should have been hydrostatically pressure tested to 1.5 times the allowable operating pressure and ultrasonically tested for metal thickness. The door should also have been totally remanufactured. In addition, all the tests should have been supervised by an authorized boiler inspector. The interior and exterior of the vessel should have been glass beaded, the exterior should have been repainted, all new piping, valves, wiring, insulation and microprocessor controls should have been installed. The anti-flood water sensing alarm should also be new. All OEM parts should have been used in the remanufacturing process and all engineering and safety upgrades made. The fact that the unit is remanufactured should be made clear in the response to bid request.
8. For a new unit, the autoclave shall have a 1 year warranty on all parts and labor and a 10 year chamber warranty against defects in materials and workmanship. Provide ASME Code construction and registered pressure vessel and three copies of ASME Form U-1 registration.
9. For a remanufactured unit, the autoclave shall have a 1 year warranty on all parts and labor and a 10 year chamber warranty against defects in materials and workmanship. Provide ASME Code construction and registered pressure vessel and three copies of ASME Form R-1 registration from a certified independent boiler inspector.
10. A full set of circuit diagrams for the control systems and schematics for the autoclave as a whole shall be supplied. Furnish drawings for installation prior to fabricating or remanufacturing the unit. The cost of shipping the unit to Gilmore Hall, Corvallis OR 97331-3906 shall be included in the bid package. The manufacturer shall install the unit and provide complete in-service training for the operators; these costs should be included
11. **The unit will be going into 105B of Gilmore Hall. Gilmore Hall does not have a freight elevator or a loading dock. Delivery will require the installer to have a lift-gate or a forklift.**



## OREGON STATE UNIVERSITY REQUEST FOR QUOTE (RFQ)

### EXHIBIT C PREVAILING WAGE RATES

In compliance with Oregon Prevailing Wage Law, the following is incorporated into this Purchase Order:

The Contractor and all subcontractors shall comply with the provisions of ORS 279C.800 through 279C.870, relative to Prevailing Wage Rates as outlined in Sections C.1 and C.2 of the General Conditions. This Purchase Order is subject to the following BOLI wage rate requirements, which are incorporated herein by Benton County

- July 1, 2015 PWR Apprenticeship Rates
- July 1, 2015 Prevailing Wage Rates for Public Works Contracts in Oregon
- July 1, 2014 Definitions of Covered Occupations for Public Works Contracts in Oregon

These BOLI wage rates are available on line at: [http://www.boli.state.or.us/BOLI/WHD/PWR/pwr\\_state.shtml](http://www.boli.state.or.us/BOLI/WHD/PWR/pwr_state.shtml)

*Remainder of page intentionally left blank.*



# OREGON STATE UNIVERSITY REQUEST FOR QUOTE (RFQ)

## EXHIBIT D INSURANCE CHECKLIST

To meet insurance requirements **all** Insurance Certificates must:

1. Be sent to **Capital Projects Contract Administration, Oregon State University, 3015 SW Western Blvd. Corvallis OR 97333**. The certificates must be "originals". We can accept PDF documents directly from your agent's office. **Faxed copies from your agent will not be accepted. Documents delivered by the contractor/vendor via fax or email will not be accepted. Documents delivered by the contractor/vendor via mail or in person, unless in a sealed envelope from your agent, will not be accepted.**
2. Contain policy numbers on the Certificate(s) **and** on any endorsements or attachments affixed to the Certificate.
3. The insurance shall contain the project name "**Gilmore Autoclave Install**" in the description on the Certificate.
4. Show the beginning and ending dates for all policies.
5. Be issued by a company licensed to do business in Oregon.
6. Contain auto liability with minimum limits of \$1,000,000 per occurrence. If coverage does not include all autos (owned, non-owned, and hired) contractor must provide written certification to OSU that only covered autos will be used for the contract. If it becomes necessary to use non-covered autos, coverage must be purchased as required by the contract and a Certificate of Insurance must be forwarded to Capital Projects Contract Administration prior to such use.
7. Contain comprehensive general liability with broad form CGL endorsement or commercial general liability (CGL) with minimum limits of \$1,000,000 per occurrence/ \$1,000,000 annual aggregate. If the "Claims Made" form of insurance is utilized, the contractor shall certify that "Tail Coverage" will be provided if such insurance is canceled, non-renewed or not replaced in like form. This is to cover claims made up to twenty-four (24) months following work acceptance. The policy shall include an occupancy clause.
8. Contain workers' compensation insurance or a signed and dated independent contractor certification statement (Form CO-3233) from the contractor.
9. Include as additional insured, by endorsement, on the liability policies, "Oregon State University, its officers, board members, agents and employees."
10. **Provide the endorsements for the general liability and auto liability w/policy # attached.**
11. **The general liability endorsement must not contain the language:** *That portion of "your work" out of which the injury or damage arises has been put to its intended use by any person or organization other than another contractor or subcontractor engaged in performing operations for a principal as part of the same project.* **If that language is included in the endorsement a completed operations with the policy number typed on it shall be provided.**
12. Give 30 days' notice of cancellation to Capital Projects Contract Administration. Any reservations must be crossed off the Certificate (i.e., "...will endeavor to mail 30 days' notice to the Certificate holder named to the left, but failure to mail such notice shall impose no obligation or liability of any kind upon the company, its agents or representatives."
13. Be signed, dated **and** contain the title of the project ; "**Gilmore Autoclave Install**".



## OREGON STATE UNIVERSITY REQUEST FOR QUOTE (RFQ)

At all times, Capital Projects Contract Administration must have a valid Certificate of Insurance (with policy dates that have not expired) as evidence that the insurance is in force. The contractor must ensure that renewal Certificates of Insurance are forwarded to Capital Project Contract Administration prior to the expiration date(s) on the Certificates. Any conflicts between this checklist and the contract requirements will be resolved in favor of the contract.

*Remainder of page intentionally left blank.*