

# SPECIFICATIONS FOR OPTICAL ATTENUATION AND ABSORPTION INSTRUMENTS ON MOBILE ASSETS AND PROFILERS (MAP)

Version 1-01
Document Control Number 1331-00007
2011-06-09

Consortium for Ocean Leadership
1201 New York Ave NW, 4<sup>th</sup> Floor, Washington DC 20005
www.OceanLeadership.org

in Cooperation with

University of California, San Diego University of Washington Woods Hole Oceanographic Institution Oregon State University Scripps Institution of Oceanography Rutgers University

#### **Document Control Sheet**

Version	Date	Description	Originator
0-01	05/10/2010	Initial Draft	Thien V. Dinh
0-02	05/24/2010	Extracted performance specs from earlier version	Rob DelCoco
0-03	08/25/2010	Reformatted per revised template; new document control number	Arthur Salwin (Noblis)
0-04	10/06/2010	Reverted to original document number. Revised definitions. Revised precision, operational depth, and wavelength range.	Arthur Salwin (Noblis)
0-05	10/26/2010	Changed "within a wavelength range" to "over a wavelength range". Referenced platform specifications in lieu of specifying operational depth range. Minor editorial fixes.	Arthur Salwin (Noblis)
0-06	11/03/2010	Revised language per approved ECR. Added references to platform specs.	Arthur Salwin (Noblis)
0-07	11/16/2010	Addressed 5-day review comments. Referenced new requirements.	Arthur Salwin (Noblis)
1-00	11/16/2010	S.E. approval – no other changes	Arthur Salwin (Noblis)
1-01	06/09/2011	Administrative update—S.E. approval	Arthur Salwin (Noblis)

Ver 1-01 1331-00007

#### **Signature Page**

This document has been reviewed and approved for release to Configuration Management.

OOI Senior Systems Engineer:

#### **Table of Contents**

1	Gen	eral	1
	1.1	Ocean Observatories Initiative (OOI) Overview	1
	1.2	Document Scope and Purpose	1
	1.3	Documents	2
	1.3.	1 Informational	2
	1.3.	2 Applicable	2
	1.4	Definitions	
	1.4.	1 Glossary and Acronyms	2
	1.4.	2 Conventions	3
2	Spe	cifications	3
	2.1	Measurement	3
	2.1.	1 Optical Absorption	3
	2.1.	2 Optical Attenuation	5
	2.2	Operational	
	2.3	Mechanical/Physical	
	2.4	Electrical	6
	2.5	Data Storage and Processing	6
	2.6	Software/Firmware	6
	2.7	Platform Interfaces	6
	2.8	Compliance	7
	2.9	Safety	7
	2.10	Shipping and Storage	7
	2.11	Identification	7
	2.12	Quality	7
3	Doc	umentation	7
4 Appendices			7

#### 1 General

#### 1.1 Ocean Observatories Initiative (OOI) Overview

Although the ocean is central to the habitability of our planet, it is largely unexplored. Biological, chemical, physical, and geological processes interact in complex ways in the ocean, at the seafloor, and at the air-sea interface. Our ability to learn more about these processes is severely limited by technical infrastructure, and developing a more fundamental scientific understanding of these relationships requires new and transformational approaches to ocean observation and experimentation.

The Ocean Observatories Initiative (OOI) will lay the foundation for future ocean science observations. OOI will enable powerful new scientific approaches by transforming the community's focus from expedition-based data gathering to persistent, controllable observations from a suite of interconnected sensors. The OOI's networked sensor grid will collect ocean and seafloor data at high sampling rates over years to decades. Researchers will make simultaneous, interdisciplinary measurements to investigate a spectrum of phenomena including episodic, short-lived events (tectonic, volcanic, oceanographic, biological, and meteorological), and more subtle, longer-term changes and emergent phenomena in ocean systems (circulation patterns, climate change, ocean acidity, and ecosystem trends).

The OOI will enable multiple scales of marine observations that are integrated into one observing system via common design elements and an overarching, interactive cyberinfrastructure. Coastal-scale assets of the OOI will expand existing observations off both U.S. coasts, creating focused, configurable observing regions. Regional cabled observing platforms will 'wire' a single region in the Northeast Pacific Ocean with a high speed optical and high power grid. Global components address planetary-scale changes via moored open-ocean buoys linked to shore via satellite. Through a unifying cyberinfrastructure, researchers will control sampling strategies of experiments deployed on one part of the system in response to remote detection of events by other parts of the system.

A more detailed discussion of the Oceans Observatories Initiative can be found in the OOI Final Network Design.

#### 1.2 Document Scope and Purpose

This document provides specifications for instruments on mobile assets and profilers that measure optical attenuation and absorption. Mobile assets for OOI include buoyancy-driven gliders and propeller-driven Autonomous Underwater Vehicles (AUVs).

Gliders are underwater autonomous vehicles that profile vertically by controlling buoyancy and moving horizontally on wings. They will be employed for two general purposes: providing horizontal context to horizontally fixed platforms, and communicating with subsurface instruments for relaying their data to shore. The buoyancy-driven, battery powered gliders change their volume by pumping to or from an oil-filled bladder; when they dive or rise; the glider's wings achieve lift allowing the glider to fly forward through the water. They can achieve speeds of about one tenth of those of the AUVs or ~25 to 35 cm s<sup>-1</sup>. At the surface, gliders acquire position information using GPS and transmit data and receive commands via satellite.

AUVs are somewhat like instrumented torpedoes, though optimized for longer life at slower speeds while carrying a sensor payload. Optimum speeds for AUVs used in oceanographic applications are near 1.7 m s<sup>-1</sup>, while maximum speeds of about 2.5 m s<sup>-1</sup> may be reached. AUVs have a high

payload capacity relative to gliders, and will carry a broad suite of sensors for interdisciplinary observations. They surface to obtain position fixes using GPS and while at the surface they also enter the OOI communications network using satellite telemetry. AUVs can run continuous missions of up to several days and are small enough to be deployed and recovered from a small boat.

Moored profilers contain a suite of sensors that are raised and lowered through the water column on a regular basis. The profiler body may travel through the water column using wire-following for deep measurements or winched technique for surface piercing measurements.

#### 1.3 Documents

#### 1.3.1 Informational

The documents listed in this section are for informational purposes only and may not have been referenced in this specification.

Consortium for Ocean Leadership, Inc. 2010, "Final Network Design", Washington, D.C. [Online] Available: <a href="http://www.oceanleadership.org/wp-content/uploads/2009/02/1101-00000">http://www.oceanleadership.org/wp-content/uploads/2009/02/1101-00000</a> FND OOI ver 2-06 Pub.pdf

#### 1.3.2 Applicable

These documents contain requirements and specifications applicable to the instrument specified. The referenced section, requirement, or specification shall be met by the instrument specified herein.

N/A

#### 1.4 Definitions

#### 1.4.1 Glossary and Acronyms

- a(λ) Absorption coefficient at wavelength λ
- **Accuracy** Closeness of the agreement between the result of a measurement and the value of the measurand (or true value of the measurement). (Taylor and Kuyatt, 1994).
- c(λ) Attenuation coefficient at wavelength λ
- **Cabled** Any OOI platform that is connected to a communications/power cable connected to shore. The platforms on the backbone cable in the Northeast Pacific are examples.
- **Coastal** For OOI, a coastal or coastal ocean site is located on the continental shelf or upper slope at a depth of 1000 m or less.
- **EIA** Electronics Industries Association
- **Instrument** A device that contains one or more sensors and a method for converting the information from the sensor into a transmittable and storable form.
- **Objective Value** The desired value of a technical parameter. This value, if provided, may be more challenging to achieve than the Threshold value. It is a goal, not a requirement, for the instrument.
- OOI Ocean Observatories Initiative
- **Open Ocean** Open ocean site is any site located at an ocean depth greater than 1000 meters or more than 500 km from shore.

- Operate Correctly performing designed functionality.
- Precision The closeness of agreement between independent measurements obtained under stipulated conditions of repeatability, generally expressed as a standard deviation (or standard uncertainty) of measurement results. Used as a measure of stability of an instrument/sensor and its capability of producing the same measurement over and over again for the same input signal (Taylor and Kuyatt, 1994).
- Resolution The smallest amount of input signal change that the instrument/sensor can detect reliably.
- PSS Practical Salinity Scale, the UNESCO Practical Salinity Scale of 1978 (PSS78).
   PSS defines salinity as a dimensionless conductivity ratio.
- **Sensor** A device that will convert a physical phenomenon into an electrical signal that can in turn be digitized through the use of an analog to digital converter. A sensor is normally housed in an instrument. Data coming from sensors is normally raw and needs to be calibrated.
- Survive Experience an event without major loss of hardware. System might experience
  loss of functionality requiring repair to return to normal mode functionality. An example of
  this is knockdown of a global mooring or loss of some part of the mooring resulting in
  the instrument descending to the bottom. Any internal memory in the instrument would
  remain accessible, but the sensors might need to be replaced to return to normal
  functionality.
- Sustain Experience an event (environmental extreme or condition) without permanent loss of normal mode functionality. System may experience reduction of functionality during event.
- Threshold Value The limiting acceptable value of a technical parameter. If this item
  does not meet the performance as specified by the threshold value, it may not be sufficient
  for inclusion in the OOI system.

#### 1.4.2 Conventions

All values contained in this document are Threshold Values unless specifically stated otherwise.

The bidder shall ignore the references in angle brackets < > at the end of each specification. They are for internal OOI use only.

#### 2 Specifications

#### 2.1 Measurement

Values provided are threshold unless otherwise stated.

#### 2.1.1 Optical Absorption

a) Measurement with unit(s)

Spectral absorption  $(a(\lambda)/m)$ 

b) Minimum Value

ABSO-001 The instrument shall measure spectral absorption over a range with a minimum value of  $0.05 \text{ a}(\lambda)/\text{m}$ .

<L2-SR-RQ-3533, L4-CG-IP-RQ-228, L4-RSN-IP-RQ-348>

c) Maximum Value

ABSO-002 The instrument shall measure spectral absorption over a range with a

maximum value of 8  $a(\lambda)/m$ .

<L2-SR-RQ-3533, L4-CG-IP-RQ-228, L4-RSN-IP-RQ-348>

d) Accuracy

ABSO-003 The instrument shall measure spectral absorption with an accuracy

within  $\pm 0.01$  a( $\lambda$ )/m of the true value. <L2-SR-RQ-3532,

L4-CG-IP-RQ-226, L4-RSN-IP-RQ-347>

e) Precision

ABSO-004 The instrument shall measure spectral absorption with a precision of

 $0.005 \text{ a}(\lambda)/\text{m}$ . <L2-SR-RQ-3742, L4-CG-IP-RQ-514,

L4-RSN-IP-RQ-567>

ABSO-005 The instrument should measure spectral absorption with a precision

of  $10^{-3}$  a( $\lambda$ )/m. This is an objective. <L2-SR-RQ-3743,

L4-CG-IP-RQ-515, L4-RSN-IP-RQ-568>

f) Resolution

Not specified.

g) Drift

Not specified.

h) Response Times

Not specified.

i) Sampling Frequency

Not specified.

j) Dependencies

Not specified.

k) Wavelength range

ABSO-006 The instrument shall measure spectral absorption over the

wavelength range of 400 nm to 720 nm. <L2-SR-RQ-3744,

L4-CG-IP-RQ-516, L4-RSN-IP-RQ-569>

ABSO-007 The instrument should measure spectral absorption over the

wavelength range of 400 nm to 750 nm. This is an objective.

<L2-SR-RQ-3745, L4-CG-IP-RQ-517, L4-RSN-IP-RQ-570>

I) Number of bands

ABSO-008 The instrument shall measure spectral absorption in no fewer than 7 bands over the wavelength range specified in 2.1.1.k. <L2-SR-RQ-3746, L4-CG-IP-RQ-518, L4-RSN-IP-RQ-571>

m) Spectral Bandwidth

ABSO-009 The instrument shall have spectral bandwidths of no more than 20 nm. <L2-SR-RQ-3747, L4-CG-IP-RQ-519, L4-RSN-IP-RQ-572>

#### 2.1.2 Optical Attenuation

a) Measurement with unit(s)

Spectral attenuation  $(c(\lambda)/m)$ 

b) Minimum Value

ATTE-001 The instrument shall measure spectral attenuation over a range with a minimum value of 0.05 c(λ)/m. <L2-SR-RQ-3537, L4-CG-IP-RQ-401, L4-RSN-IP-RQ-354>

c) Maximum Value

ATTE-002 The instrument shall measure spectral attenuation over a range with a maximum value of 15 c(λ)/m.
<L2-SR-RQ-3537, L4-CG-IP-RQ-401, L4-RSN-IP-RQ-354>

d) Accuracy

ATTE-003 The instrument shall measure spectral attenuation with an accuracy within  $\pm 0.01$  c( $\lambda$ )/m of the true value. <L2-SR-RQ-3536, L4-CG-IP-RQ-399, L4-RSN-IP-RQ-353>

e) Precision

ATTE-004 The instrument shall measure spectral attenuation with a precision of 0.005 c(λ)/m. <L2-SR-RQ-3748, L4-CG-IP-RQ-520, L4-RSN-IP-RQ-573>

ATTE-005 The instrument should measure spectral attenuation with a precision of  $10^{-3}$  c( $\lambda$ )/m. This is an objective. <L2-SR-RQ-3749, L4-CG-IP-RQ-521, L4-RSN-IP-RQ-574>

f) Resolution

Not specified.

g) Drift

Not specified.

h) Response Times

Not specified.

i) Sampling Frequency

Not specified.

j) Dependencies

Not specified.

k) Wavelength range

ATTE-006 The instrument shall measure spectral attenuation over the

wavelength range of 400 nm to 720 nm. <L2-SR-RQ-3750,

L4-CG-IP-RQ-522, L4-RSN-IP-RQ-575>

ATTE-007 The instrument should measure spectral attenuation over the

wavelength range of 400 nm to 750 nm. This is an objective.

<L2-SR-RQ-3751, L4-CG-IP-RQ-523, L4-RSN-IP-RQ-576>

I) Number of bands

ATTE-008 The instrument shall measure spectral attenuation in no fewer than 7

bands over the wavelength range specified in 2.1.2.k.

<L2-SR-RQ-3752, L4-CG-IP-RQ-524, L4-RSN-IP-RQ-577>

m) Spectral Bandwidth

ATTE-009 The instrument shall have spectral bandwidths of no more than

20 nm. <L2-SR-RQ-3753, L4-CG-IP-RQ-525, L4-RSN-IP-RQ-578>

2.2 Operational

See platform specifications.

2.3 Mechanical/Physical

See platform specifications.

2.4 Electrical

See platform specifications.

2.5 Data Storage and Processing

See platform specifications.

2.6 Software/Firmware

See platform specifications.

2.7 Platform Interfaces

See platform specifications.

2.8 Compliance

See platform specifications.

2.9 Safety

See platform specifications.

2.10 Shipping and Storage

See platform specifications.

2.11 Identification

See platform specifications.

2.12 Quality

See platform specifications.

#### 3 Documentation

See the RFP for documentation that the vendor shall be required to supply.

#### 4 Appendices

None