

REQUEST FOR PROPOSALS (RFP)
Energy Dispersion X-ray Spectrometers
For
Portland State University
(RFP No. 21051)

ADDENDUM Number 1

REQUEST FOR PROPOSAL: ADDENDUM Number 1

RFP NO: 21051

TITLE: Energy Dispersion X-ray Spectrometers

DATE RFP ISSUED: February 4, 2013

DATE ADDENDUM 1 ISSUED: February 7, 2013

CONTACT PERSON: Paul L. Thomas
PHONE/FAX: 503-725-9841/503-725-7873
Email: contract@pdx.edu

PURPOSE OF THIS ADDENDUM 1:

The purpose of this Addendum 1 is to respond to those questions or requests for clarification/change (or protest) submitted by potential Proposers pursuant to Section 1, Paragraph 3 of the RFP. The questions or requests for clarification/change (or protest) are in italics. PSU's response is underlined.

VENDOR QUESTION:

On page 12 of the bid spec, item 2b (sic, the correct citation is: Section 2: Scope of Work, Minimum Mandatory Requirements, A. Mandatory Specifications, item 2.e.) states "The detector will have a motorized slide".

The potential Proposer does not believe a motorized slide is necessary for SEM operation.

PSU's RESPONSE:

PSU has considered the potential Proposer's request for change of Section 2: Scope of Work, Minimum Mandatory Requirements, A. Mandatory Specifications, item 2.e., and PSU hereby rejects the potential Proposer's request for change.

Section 2: Scope of Work, Minimum Mandatory Requirements, A. Mandatory Specifications, item 2.e. shall remain as originally stated, "The detector will have a motorized slide."

NOTICE: PSU assumes no liability for inadvertent errors or mistakes in this document. PSU has answered all vendor questions to the best of our ability with the information we have readily available at the time of issuance of this document.