



**Construction Contract Services**  
Oregon State University, 3015 SW Western Blvd, Corvallis, Oregon 97333  
T 541-737-7342 | F 541-737-4810

### ADDENDUM

**SOLICITATION #:** BD164030P

**SOLICITATION NAME:** WIEGAND STERLIZER INSTALL

**ADDENDUM #:** 2

**DUE DATE AND TIME:** APRIL 24, 2013 3:00 PM PST

**DATE:** APRIL 19, 2013

**PROCUREMENT ANALYST:** BROOKE DAVISON

**The solicitation named above is hereby modified as follows:**

1. Informal Request for Proposal, Page 2, Section 3.01 titled "Required Specifications", is modified to add the following bullet point:
  - a. "D. Demolition and removal of existing autoclave required.
    - 1) No modifications or damage to existing drain system during demolition allowed."
2. Informal Request for Proposal, Page 2, Section 3.01 titled "Required Specifications", is modified to add the following instruction for bullet point B:
  - a. "1) Installation must connect to existing drain without modifications to building drain system."
  - b. "2) Existing Drain is approximately 8"-12" above floor."

**The following questions were received with regard to the solicitation named above. OSU has provided answers below to each question, but the RFP or contract documents have not been modified as a result.**

1. **What is the steam pressure available at Weigand?**  
**Answer: ~60psi**
2. **Could a cabinet with two sides be an option?**  
**Answer – No**
3. **What floor is sterilizer to be installed on?**  
**Answer – 1<sup>st</sup> Floor**
4. **Is there a dock available for building?**  
**Answer – Yes**
5. **Is there a clear path from dock to final location of sterilizer?**  
**Answer – Yes**

Entities are not required to return addendums. However, entities are responsible to make themselves aware of, obtain and incorporate any changes made in any addendums issued into their final offer. Failure to do so may, in effect, make the offer non-responsive, which may cause the offer to be rejected.