



Procurement and Contract Services (PaCS)
Oregon State University, 644 SW 13th St, Corvallis, Oregon 97333-4238
T 541-737-4261 | F 541-737-2170 | <http://pacs.oregonstate.edu/>

ADDENDUM

SOLICITATION #: ML164886Q

SOLICITATION NAME: DIAGNOSTIC ULTRASOUND SYSTEM

ADDENDUM #: 1

DUE DATE AND TIME: JUNE 12TH, 2013 @4:00PM

DATE: JUNE 12TH

PROCUREMENT ANALYST: MARK LESSEL

The solicitation named above is hereby modified as follows:

1. Replaced RFQ# ML164886Q form in its entirety with the following revised RFQ# ML164886Q-Rev1 form. (Changes are highlighted)
2. Change RFQ DUE DATE: from June 12th, 2013 @ 4:00 PM to June 18th, 2013 @4:00PM

Entities are not required to return addendums. However, entities are responsible to make themselves aware of, obtain and incorporate any changes made in any addendums issued into their final offer. Failure to do so may, in effect, make the offer non-responsive, which may cause the offer to be rejected.



**OREGON STATE UNIVERSITY
REQUEST FOR QUOTE (RFQ)**

RFQ #		ML164886Q-REV1		ISSUE DATE:	June 6, 2013
DELIVER TO:		REQUESTED BY / RETURN QUOTE TO:			
DEPARTMENT:	Veterinary Medicine	NAME:	Mark Lessel		
ADDRESS:	Oregon State University	E-MAIL:	Mark.lessel@oregonstate.edu		
CITY, STATE ZIP:	Corvallis OR, 97331	TELEPHONE:	541-737-3667		
REQUIRED DELIVERY DATE:	6-28-2013	FAX:	541-737-2170		

ITEM	DESCRIPTION	QTY	UNIT	UNIT PRICE	TOTAL PRICE
1	Diagnostic Ultrasound System, Fully-digital ultrasound system which is fully portable a laptop-style ultrasound system	1	ea		
	Portable, Laptop-Style ultrasound system with at least 15" LCD monitor.				
	Color Doppler, Pulse subtraction THI, Pulsed wave Doppler and dual/twin view. Contrast harmonic imaging with real-time dual view. Additional functions such as 3D volume imaging, elastography and Omni-directional M-mode.				
	Mountable to a pole-type cart, transportable. The cart has to be equipped with slots for at least 2-3 adapters for the ultrasound probes.				
	Probe adaptor for Transducers				
	ECG waveform display				
	Standard battery and AC power adapter. Battery pack ability to scanning for at least 30 minutes without additional external electricity support				
	DICOM connectable, ability to send pictures to picture-archival storage server				
	The unit has to come with at least 3 broadband transducers, which should include one linear probe in the range of approximately 5-20 MHz, one micro convex probe of approximately 2-15 MHz and one micro-convex probe of approximately 1-8 MHz. The linear probe has to be covering the high frequency spectrum of up to at least 15 MHz				
	The ultrasound system has to weigh less than 25lb				
	Additional wireless function to send images is preferred. Another preferred function is to retrieve images from the PACS and being able to view them on the ultrasound machine in real-time while an ultrasound study is being performed				

Vendor Name: _____

Signature: _____



**OREGON STATE UNIVERSITY
REQUEST FOR QUOTE (RFQ)**

Delivery is f.o.b. destination, prepaid and allowed. Shipping, freight and handling must be included in quoted prices. Additional costs for such are disallowed.		TOTAL	
DELIVERY TIME AFTER RECEIPT OF ORDER:		PRICES VALID THROUGH:	
SPECIAL INSTRUCTIONS: 1. Unless otherwise specified, all items quoted are to be new, unused and not remanufactured in any way. 2. Brand names are for the purpose of describing and establishing the characteristics desired and are not intended to limit or restrict competition. Quoters may submit quotes for substantially equivalent products unless the RFQ provides that a specific brand is necessary because of compatibility requirements, etc. All such brand substitutions shall be subject to approval by OSU. 3. Quoters must clearly identify all products quoted. Brand name and model or number must be shown. 4. Only documents issued as addenda by OSU serve to change the RFQ in any way. 5. OSU reserves the right to make the award by item, partial or whole lots, groups of items or entire quote, whichever is in the best interest of OSU. 6. OSU may reject any Quote not in compliance with the RFQ, attachments, and addenda, or if it is in the best interest of OSU.	VENDOR INFORMATION:		
	COMPANY:		
	ADDRESS:		
	CITY, STATE, ZIP:		
	CONTACT NAME:		
	E-MAIL:		
	TELEPHONE:		
	FAX:		
	VENDOR SIGNATURE: <i>By signature below the undersigned certifies that they are authorized to act on behalf of the quoter and will comply with all aspects of the quote herein.</i>		
	SIGNATURE:		
NAME/TITLE:			

This procurement is subject to the indicated Oregon State University Standard Terms and Conditions for: Goods Services Purchase Order Construction Software. The indicated terms and conditions may be viewed at <http://pacs.oregonstate.edu/terms-and-conditions>