



ADDENDUM

SOLICITATION NO.: JD173603P

SOLICITATION NAME: REPLACEMENT OF THE ⁶⁰Co SOURCES FOR GAMMA CELL GC-220 IRRADIATOR AT OREGON STATE UNIVERSITY

ADDENDUM NO.: 1

DUE DATE AND TIME: DECEMBER 16TH, 3:00PM PST

DATE: 12-4-2014

PROCUREMENT ANALYST: JOSHUA DODSON

1. Page 3, Section 3.01 titled "Statement of Work" Paragraph A, **REMOVE** last sentence "The nominal total activity should be a minimum of approximately 7,000 Ci and a maximum of 20,000 Ci" and **REPLACE WITH** "The nominal total activity should be a minimum of 7,000 Ci and a maximum of 10,000 Ci. Vendor is encouraged to provide price points for additional Ci Levels above the minimum but will only be scored in Exhibit D based on their rate for the replacement of sources up to the 7000 Ci Level."
2. Page 3, Section 3.01 title "Statement of Work" Paragraph B, after last sentence add "If sources are not yet manufactured, participant does not need to provide serial number, activity, and assay date with their submission. However, a successful participant that plans to manufacturer sources after award, will need to provide this data to OSU as soon as it becomes available."
3. Page 4, Section 5.02 titled "Required Submittals" Paragraph B, after last sentence add "OSU only requires that existing sources be removed and disposed of if they're removal is necessary for the installation of new sources that will meet the minimum 7000Ci level requirement. The participant will only be scored in Exhibit D based on their rate for the replacement of sources required for new source installation However, participants are encouraged to provide pricing for the removal of the remaining sources separately in their proposal package."
4. Page 4, Section 5.02 titled "Required Submittals" Paragraph D, is modified to add the following after the last sentence. "If the participant has not had any violations during the timeframe listed above, a statement can be provided in place of the correspondence requirement asserting that no violations have occurred."
5. Page 4 Section 5.02 title "Required Submittals" Paragraph H is modified to add the following after the last sentence. "Participants must provide at least 1 Reference in order to meet the Exhibit C requirement. References should not include any confidential sources. Participants are encouraged to include references for work that is closely related to the scope of work laid out in this proposal."

Answers to questions not covered in the solicitation changes above are on page 2 of this Addendum.

The following questions were received with regard to the solicitation named above. OSU has provided answers below to each question, but the RFP or contract documents have not been modified as a result.

Q: When would OSU like this work to be performed?

A: As soon as possible.

Q: Does the Gamma Cell 220 unit require additional repairs?

A: No, the unit is functional. OSU just needs to have the current ^{60}Co sources replaced to bring the Ci level up to a minimum of 7000Ci.

Q: Is the building where the unit resides 2 stories high? If so, what is the plan for minimizing upward radiation exposure during the source transfer?

A: The building is only 1 story. OSU would prefer that the work is completed within the room where the unit currently resides.

Entities are not required to return addendums with their offers but are responsible to make themselves aware of, obtain and incorporate into their final offer any information contained in addendums. Failure to do so may make the offer non-responsive and cause it to be rejected.