



**OREGON STATE UNIVERSITY
REQUEST FOR QUOTE (RFQ)**

		ISSUE DATE:	June 8, 2012		
RFQ #	DL160021Q	RFQ DUE DATE:	June 13, 2012, 2 PM		
DELIVER TO:		REQUESTED BY / RETURN QUOTE TO:			
DEPARTMENT:	Food Science	NAME:	Debora Lauer		
ADDRESS:	OSU Wiegand Hall	E-MAIL:	Debora.Lauer@oregonstate.edu		
CITY, STATE ZIP:	Corvallis, OR 97331	TELEPHONE:	541-737-7343		
REQUIRED DELIVERY DATE:	July 13, 2012	FAX:	541-737-2170		
ITEM	DESCRIPTION	QTY	UNIT	UNIT PRICE	TOTAL PRICE
1	CTC AOC 5000-xt FOR STATIC HEADSPACE, 120 cm RAIL	1	Ea		
2	2 Drawer Sample Tray Stack, Peltier cooled	1	Ea		
3	Mount Kit for AOC-5000 on GC-2010 (PAL MR S2010)	1	Ea		
4	CYCLE COMP S/W, SIL5000/AOC-5000	1	Ea		
5	CTC AUTO SAMPLER CONTROL, FIRST OPTION	1	Ea		
6	AOC 5000 SPME OPTION, PAL SPME-Kit-USA	1	Ea		
7	SPME FIBER CONDITIONING STATION	1	Ea		
8	GCsolution Control S/W for AOC-5000	1	Ea		
9	SMPL TRAY 32 VIALS 10/20ML W/TRAYHOLDER 114MM	1	Ea		
(See page 2 for more detailed specifications)					
SHIPPING MUST BE INCLUDED IN THE QUOTE					
Delivery is f.o.b. destination, prepaid and allowed. Shipping, freight and handling must be included in quoted prices. Additional costs for such are disallowed.				TOTAL	
DELIVERY TIME AFTER RECEIPT OF ORDER:			PRICES VALID THROUGH:		
SPECIAL INSTRUCTIONS:		VENDOR INFORMATION:			
<p>1. Unless otherwise specified, all items quoted are to be new, unused and not remanufactured in any way.</p> <p>2. Brand names are for the purpose of describing and establishing the characteristics desired and are not intended to limit or restrict competition. Quoters may submit quotes for substantially equivalent products unless the RFQ provides that a specific brand is necessary because of compatibility requirements, etc. All such brand substitutions shall be subject to approval by OSU.</p> <p>3. Quoters must clearly identify all products quoted. Brand name and model or number must be shown.</p> <p>4. Only documents issued as addenda by OSU serve to change the RFQ in any way.</p> <p>5. OSU reserves the right to make the award by item, partial or whole lots, groups of items or entire quote, whichever is in the best interest of OSU.</p> <p>6. OSU may reject any Quote not in compliance with the RFQ, attachments, and addenda, or if it is in the best interest of OSU.</p>		COMPANY:			
		ADDRESS:			
		CITY, STATE, ZIP:			
		CONTACT NAME:			
		E-MAIL:			
		TELEPHONE:			
FAX:					
		VENDOR SIGNATURE:			
		<i>By signature below the undersigned certifies that they are authorized to act on behalf of the quoter and will comply with all aspects of the quote herein.</i>			
		SIGNATURE:			
		NAME/TITLE:			

This procurement is subject to the indicated Oregon State University Standard Terms and Conditions for: Goods Services Purchase Order Construction Software. The indicated terms and conditions may be viewed at <http://pacs.oregonstate.edu/terms-and-conditions>



OREGON STATE UNIVERSITY REQUEST FOR QUOTE (RFQ)

Autosampler Accessory

Must include an autosampler for liquid and static headspace injections (SPME)

Power supply – Must have an input voltage of 100-240V, 120W, 50/60Hz and come with the necessary cable. 1 control terminal & cable

Injection unit – Must have an injection unit that can use both liquid injectors and headspace fibers.

Dimensions – X-axis or length of the unit must be 1200mm in order to be used for MDGC and access Peltier cooling unit

Incubator – Must be able to accommodate 20mL vials, shake/agitate the vials up to speeds of 250 rpm and provide heating temperatures of up to 200°C.

Sample trays – Must include 1 sample tray for 2mL vials (holds 98 vials), 1 mL vials (holds 78 vials) and 10/20mL vials (holds 32 vials). 2 sample tray holder, 114 mm

Syringes – Must contain 1 syringe adapter 10uL with plunger holder and contain a 1.5mL syringe heater that can reach a maximum temp of 150°C. Must also include 2 headspace syringes with 2.5mL safety guard.

Peltier cooler – Must include a tray that can cool 20mL vials, able to achieve cooling temperatures as low as 0°C.

Software – Include the latest version of cycle composer software that is compatible with the chosen GC & GCMS software.

SPME fiber conditioning station – Must